

simplifying research

qPCRBIO Probe 1-Step Virus Detect Separate-ROX www.pcrbio.com

## **Product description**

qPCRBIO Probe 1-Step Virus Detect is designed for highly sensitive 1-step RT-qPCR-based detection of viral RNA sequences. The kit has been optimised with a high concentration 4x mix, enabling greater sample input and increased sensitivity, even when small volume reactions are used.

qPCRBIO Probe 1-Step Virus Detect is engineered for use with a wide range of probe technologies including TaqMan®, Scorpions® and molecular beacon probes. The kit is compatible with multiplexing assays and can be used to detect viral RNA sequences over a broad range of template concentrations, down to 4 copies per reaction (0.8 copies per μL).

The kit includes the thermostable UltraScript Reverse Transcriptase which is blended with an advanced RNase inhibitor to prevent degradation of RNA by contaminating RNase. Combined with antibody-mediated hot start technology and smart screen buffer chemistry, qPCRBIO Probe 1-Step Virus Detect enables robust, reproducible and high throughput detection of RNA viruses.

### Detection of SARS-CoV-2

qPCRBIO Probe 1-Step Virus Detect has been validated for qualitative detection of SARS-CoV-2 nucleic acid using the Charité (Berlin, Germany) recommended primer-probe sequences (RdRp and E genes)<sup>1</sup>, and CDC (Atlanta, USA) primer-probe sequences (N gene)<sup>2</sup>. For further information please email technical@pcrbio.com.

| Component                                         | 200 rxns  | 600 rxns  | 1000 rxns |
|---------------------------------------------------|-----------|-----------|-----------|
| 4x qPCRBIO Probe<br>1-Step Virus Detect<br>No-ROX | 1 x lmL   | 3 x 1mL   | 1 x 5mL   |
| 50µM ROX Additive                                 | 1 x 200µL | 1 x 200µL | 1 x 200µL |
| 20x UltraScript RTase<br>(with RNase inhibitor)   | 1 x 200µL | 1 x 600µL | l x lmL   |

## Shipping and storage

On arrival the kit should be stored between -30°C and -15°C. Avoid prolonged exposure to light. If stored correctly the kit will retain full activity for 12 months. The kit can go through 30 freeze/thaw cycles with no loss of activity.

# Limitations of product use

For research use only. This product alone does not provide any diagnostic result.

# Technical support

Help and support is available on our website at <a href="https://pcrbio.com/resources/">https://pcrbio.com/resources/</a> including answers to frequently asked technical questions. For technical support and troubleshooting please email technical@pcrbio.com with the following information:

- Amplicon size
- Reaction setup
- Cycling conditions
- Screen grabs of amplification traces and melting profile



T 041 250 53 57 | mail@witec.ch

<sup>&</sup>lt;sup>1</sup> Diagnostic detection of 2019-nCoV by real-time RT-PCR (https://www.who.int/docs/default-source/coronaviruse/protocol-v2-1.pdf)

<sup>&</sup>lt;sup>2</sup> 2019-Novel Coronavirus (2019-nCoV) Real-time rRT-PCR Panel Primers and Probes (https://www.cdc.gov/coronavirus/2019-ncov/ downloads/rt-pcr-panel-primer-probes.pdf)

#### Important considerations

Instrument compatibility: Different real-time PCR instruments require different levels of ROX passive reference. Generally, modern instruments do not need passive reference but include the option to use it for normalisation. Please use our qPCRBIO Selection Tool to determine which ROX concentration your instrument requires (https://pcrbio.com/resources/qpcr-selection-tool/).

ROX additive protocol: The 50µM ROX Additive supplied is formulated to be added directly to the tube or bottle of 4x qPCRBIO mix supplied. Once the ROX is added, the reagent may be used straight away or stored between -30°C and -15°C for future use. Please use the following 2 charts to add the correct amount of ROX for your instrument. Vortex thoroughly after ROX addition.

| Hi-ROX instruments                          | Reagent volume   | Final concentration                     | Reaction concentration |
|---------------------------------------------|------------------|-----------------------------------------|------------------------|
| 4x qPCRBIO Probe 1-Step Virus Detect No-ROX |                  | 2x                                      | 1x                     |
| 50μM ROX Additive                           | 20.0µL           | lμM                                     | 500nM                  |
| Lo-ROX instruments                          | Reagent volume   | Final concentration                     | Reaction concentration |
|                                             | •                | 7 mar concentration                     | 1.                     |
| 4x qPCRBIO Probe 1-Step Virus Detect No-ROX | I.UIIL           | ZX                                      | 1 X                    |
|                                             | **************** | · · · • · · · · · · · · · · · · · · · · | *                      |

Template: The kit can be used with RNA extracted by most commercial kits, provided the amount and quality of template RNA are within an acceptable range. Addition of sample as 2 to  $5\mu$ L volumes will improve assay precision.  $5\mu$ L of swab extract is recommended for SARS-CoV-2 diagnostic assays.

#### Reaction setup

- 1. Before starting, briefly vortex 4x qPCRBIO Probe 1-Step Virus Detect mix
- 2. Prepare a master mix based on the following table. We also recommend setting up a no-RTase control:

| Reagent                                    | 20µL reaction           | Final conc. | Notes                                                                                               |
|--------------------------------------------|-------------------------|-------------|-----------------------------------------------------------------------------------------------------|
| 4x PCRBIO Probe 1-Step Virus<br>Detect mix | 5μL                     | lx          |                                                                                                     |
| Forward primer (10µM)                      | 1-2µL                   | 400nM-1μM   |                                                                                                     |
| Reverse primer (10µM)                      | 1-2µL                   | 400nM-1μM   |                                                                                                     |
| Probe (10µM)                               | 0.25-1μL                | 125-500nM   |                                                                                                     |
| 20x UltraScript RTase                      | 1μL                     | 1x          |                                                                                                     |
| RNA template                               | 2-5µL                   | Variable    | 4 to 1x10 <sup>8</sup> viral copies per reaction.<br>See above for further template considerations. |
| PCR grade dH <sub>2</sub> O                | Up to 20µL final volume |             |                                                                                                     |

3. Program the instrument using the following conditions, acquiring data on the appropriate channel:

| Cycles           | Temperature<br>General | Temperature<br>SARS-CoV-2 Detection | Time                                         | Notes                                                                      |
|------------------|------------------------|-------------------------------------|----------------------------------------------|----------------------------------------------------------------------------|
| 1                | 45°C to 55°C           | 55°C                                | 5 minutes singleplex<br>10 minutes multiplex | Reverse transcription                                                      |
| 1                | 95°C                   | 95°C                                | 3 minutes                                    | Polymerase activation and RTase inactivation                               |
| 50               | 95°C<br>55°C to 65°C   | 95°C<br>58°C                        | 15 seconds<br>30 seconds                     | Denaturation<br>Anneal/Extension                                           |
| Melt<br>analysis | Refer to instrum       |                                     |                                              | Optional melt profile analysis, available for<br>hybridisation probes only |